

# IHS Inclinometer


Designed by Craig J. Della Valle, MD

*Helps to accurately predetermine angles for acetabular cup positioning and insertion*  
*With a measuring scale from 0 to 45°, the inclinometer may be used on the reamer shaft, the trial cup shaft and the cup impactor shaft.*

Designed to help allow the surgeon to consistently and quickly achieve the desired component position during each step of acetabular preparation and component positioning: acetabular reaming, trial component positioning, and actual component insertion.



Bottom Profile with Magnets



PRODUCT NO:

**1326**

Dimensions: 4" x 2" (10,2 cm x 5,1 cm)



**\* Accuracy Statement for devices with a measuring function:**

Devices with a measuring function are considered non-critical, non-calibrated, and are intended for reference-only measurements. While it provides a helpful visual guide for relative measurements, it is not calibrated to specific precision standards. Measurements taken with this device should not be considered accurate enough for critical applications requiring strict dimensional tolerances. General manufacturing tolerances of  $\pm 1\%$  at each graduation, or 2% of the full scale apply, but no conformance qualifications are provided or implied.

[www.innomed.net](http://www.innomed.net)



ISO 13485:2016

**FREE TRIAL ON MOST INSTRUMENTS**



103 Estus Drive, Savannah, GA 31404  
info@innomed.net

912.236.0000 Phone  
912.236.7766 Fax

Innomed-Europe Tel. +41 41 740 67 74  
info@innomed-europe.com

**INNOMED**

**1.800.548.2362**